Tissue Processor YD-14P1.8

Specification			
1	**Quantity of Tank**	14	10 reagent tank and 4 paraffin tank
2	**Basket**	2	Can run two baskets together, and execute two different programs.
3	**Capacity of single tray**	1500 ml	Can load 300 samples or 120pcs cassettes(240pcs cassettes if run 2 baskets simultaneously)
4	**Temperature range for paraffin**	50℃~89℃	Increments 1℃
5	**Setting range time for each step**	From 1 min. to 99 h 59 min.	
6	**Store program**	6 sets	
7	**Dripping stay time**	30s	Reduce reagent pollution
8	**Up and down agitation**	4 times/ each 5 min.	
9	**Setting delaying process time**	99 h 99 min.	
10	**Dimension**	140×50×55cm	

Function		
1	**Two tissue baskets running together**	Meet different customers` requirements. It can run two baskets together while tissue quantity large. Otherwise run one basket to deal with few tissue.
2	**Run two different programs simultaneity**	While two baskets running together, it can execute the same or different programs. So the big and small tissue, or hard and soft tissue can be dehydrate separately, to improve the dehydration quality.
3	**With ion filter**	Ion filter can clean out the baneful gas completely, to make sure users safe.
4	**Colorized touch screen**	Display all running parameters, easy to watch and convenient to operation.
5	**With delay process time function**	While in automatic running, if need add the running time in any station, use this function easy to modify the program.
6	**Movement parts**	Movements parts choose bearing linear guide, high precision, small resistance, steady running, strong abradability.
8	**Heating paraffin station**	Precision temperature control

Standard Accessory
1 Power Wire
2 pcs Fuse

See Price list for more detail !

Jinhua YIDI Medical Appliance CO.,LTD
Zhuma industrial park,jinhua city, zhejiang ,china
Telephone: 86 579 82472777 Fax:86 579 82472470
www.zjyidi.com